

CDLC1102

USED WITH:
ENAMEL RATER END TESTING CUP

CHANGE PARTS

9756
9757
9758
9759



9756



9757



9758



9759

LATEST VERSION(S) MAY VARY SLIGHTLY FROM PHOTO(S)

FOR DIRECT TECHNICAL ASSISTANCE USE FAX # (860) 496-0255

PRODUCT OVERVIEW

These Change Parts are used with enamel raters to rate the enamel coating on an unseamed end. The Change Parts are manufactured by Altek to specific customer requirements.

9756 DESCRIPTION

The 9756 Change Parts hold a round end in place as it's enamel coating is being rated. The 9756 consists of a plexiglas end adapter assembly which fits onto the end holder cup, holding the end during testing.

9757 DESCRIPTION

The 9757 Change Parts hold a single-radius rectangular end in place as it's enamel coating is being rated. The 9757 consists of a plexiglas end adapter assembly which fits onto the end holder cup, holding the end during testing.

9758 DESCRIPTION

The 9758 Change Parts hold a multiple-radius rectangular end in place as it's enamel coating is being rated. The 9758 consists of a plexiglas end adapter assembly which fits onto the end holder cup, holding the end during testing.

9759 DESCRIPTION

The 9759 Change Parts hold an oval end in place as it's enamel coating is being rated. The 9759 consists of a plexiglas end adapter assembly which fits onto the end holder cup, holding the end during testing.

SPECIFICATIONS

The following document describes the specifications for these Change Parts:

CDXC1101 - Change Parts Specifications

9756 ORDERING

The 9756 is ordered by specifying 9756 on the appropriate Customer Order Guide and supplying several samples to Altek. Samples MUST be provided to ensure proper manufacture of the 9756 Change Parts.

9757 ORDERING

The 9757 is ordered by specifying 9757 on the appropriate Customer Order Guide and supplying several samples to Altek. Samples MUST be provided to ensure proper manufacture of the 9757 Change Parts.

9758 ORDERING

The 9758 is ordered by specifying 9758 on the appropriate Customer Order Guide and supplying several samples to Altek. Samples MUST be provided to ensure proper manufacture of the 9758 Change Parts.

9759 ORDERING

The 9759 is ordered by specifying 9759 on the appropriate Customer Order Guide and supplying several samples to Altek. Samples MUST be provided to ensure proper manufacture of the 9759 Change Parts.

TECHNICAL SUPPORT

Altek provides comprehensive technical support to its customers worldwide. For assistance, contact Altek and ask for the Engineering department.

REPAIR

Altek customers can obtain a return authorization from Altek and ship a faulty unit in for repair, if necessary. In addition, replacement parts for all models are available from Altek.

WARRANTY

All Altek products are built to the highest standards of quality and reliability. This Altek product is backed by a full 90-day warranty.

FURTHER INFORMATION

See the accompanying specifications for additional information concerning this product. If specifications were not included with this flier, please call Altek; they will be mailed or faxed as quickly as possible

CUSTOM DESIGN.

Altek can custom design inspection and test equipment to meet specific customer requirements. Custom Change Parts can also be designed and manufactured for most Altek products; contact Altek for further information.

A NOTE ON SAMPLES

Samples sent to Altek as a requirement for the manufacture of Change Parts must be submitted in their "as tested" form. For instance, if an end will be tested unseamed on a piece of test equipment, Altek must receive ends that are unseamed in order to properly manufacture the Change Parts. Contact Altek for further information regarding sample requirements for Change Parts manufacture.

For more information, contact Altek between 8:00 am and 5:00 pm ET, Monday thru Friday. The Altek company address and phone numbers can be found on the front of this flier.

ORDERING INFORMATION

To purchase this product:

This flier is supplemented by additional literature in the form of a specification sheet.

These Change Parts are not required for an enamel rater if only cans are being rated.

At least one of these Change Parts sets is required for operation of an Altek enamel rater end testing cup.

Customer samples are REQUIRED in order to ensure proper manufacture and operation of the Change Parts.

These Change Parts are ordered by specifying the required information (as described above) on the appropriate Customer Order Guide.

Thank you for your interest in TECH • products from Altek!

CDXC1101.002

C11 SPECIFICATIONS

ALL SPECIFICATIONS ARE SUBJECT TO CHANGE WITHOUT NOTICE

CHANGE PARTS: 9756, 9757, 9758, and 9759

USED WITH MODELS: 9502, 9524A, 9526A

EFFECTIVE DATE: MARCH 4, 1998

SAMPLE TYPE: BEVERAGE AND FOOD/UTILITY ENDS

NOTE: A SEPARATE SET OF CHANGE PARTS MUST BE MANUFACTURED FOR EACH SPECIFIC DIAMETER OR CROSS-SECTION OF SAMPLE.

NOTE: A DIAMETER ENVELOPE LIMIT MEANS THAT THE SAMPLE MUST FIT COMPLETELY WITHIN A CIRCLE OF THE GIVEN DIAMETER, REGARDLESS OF THE SPECIFIC SHAPE OF THE SAMPLE.

9756: For round, unseamed metal beverage and food/utility ends

End Diameter 200 - 502
(2" to 5 2/16")

9757: For single-radius rectangular, unseamed metal food/utility ends

End L x W 502 dia envelope max
(5 2/16")

9758: For multiple-radius rectangular, unseamed metal food/utility ends

End L x W 502 dia envelope max
(5 2/16")

9759: For oval, unseamed metal food/utility ends

End L x W 502 dia envelope max
(5 2/16")