

CDLC1702

**CHANGE PARTS**

**9818**

**USED WITH:  
END BUCKLING TESTERS**



**9818**

LATEST VERSION(S) MAY VARY SLIGHTLY FROM PHOTO(S)

## PRODUCT OVERVIEW

These Change Parts are used with end buckle testers to hold a free end in place as it is being pressurized with air. The Change Parts are manufactured by Altek to specific customer requirements.

### 9818 DESCRIPTION

The 9818 Change Parts hold an unseamed end in place as it is being pressurized. The 9818 consists of an end holder base with a molded synthetic gasket on which the free end rests, a clamping ring which is placed on top of the end and a pair of jaws which clamp the sample between the clamping ring and end holder during testing.

The 9818 is also available with a clamping ring that restrains the counter-sink area and prevents the end from buckling, necessary for the testing of Tab Blow-out or Vent failures.

## SPECIFICATION / ORDER GUIDE

The following document describes the specifications for these Change Parts:

CDXC17## - Change Parts Specifications

### 9818 ORDERING

The 9818 is ordered by specifying 9818 and supplying several samples to Altek. Samples MUST be provided to ensure proper manufacture of the 9818 Change Parts. Detail Drawings are also helpful.

## TECHNICAL SUPPORT

Altek provides comprehensive technical support to its customers worldwide. For assistance, contact Altek and ask for the Engineering department.

### REPAIR

Altek customers can obtain a return authorization from Altek and ship a faulty unit in for repair, if necessary. In addition, replacement parts for all models are available from Altek.

### WARRANTY

All Altek products are built to the highest standards of quality and reliability. This Altek product is backed by a full 90-day warranty.

### FURTHER INFORMATION

See the accompanying specifications for additional information concerning this product. If specifications were not included with this flier, please call Altek; they will be mailed or faxed as quickly as possible

### CUSTOM DESIGN.

Altek can custom design inspection and test equipment to meet specific customer requirements. Custom Change Parts can also be designed and manufactured for most Altek products; contact Altek for further information.

### A NOTE ON SAMPLES

Samples sent to Altek as a requirement for the manufacture of Change Parts must be submitted in their "as tested" form. For instance, if an end will be tested unseamed on a piece of test equipment, Altek must receive ends that are unseamed in order to properly manufacture the Change Parts. Contact Altek for further information regarding sample requirements for Change Parts manufacture.

For more information, contact Altek between 8:00 am and 5:00 pm ET, Monday thru Friday. The Altek company address and phone numbers can be found on the front of this flier.

## ORDERING INFORMATION

To purchase this product:

This flier is supplemented by additional literature in the form of a specification sheet.

At least one of these Change Parts sets is required for operation of an Altek buckle tester.

Customer samples are REQUIRED (and detail drawings are helpful) in order to ensure proper manufacture and operation of these Change Parts.

*Thank you for your interest in TECH • products from Altek!*

CDXC1701.007

C17 SPECIFICATIONS

ALL SPECIFICATIONS ARE SUBJECT TO CHANGE WITHOUT NOTICE

CHANGE PARTS: ..... 9818

USED WITH MODELS: ..... 9009C4, 9009C4F, 9009H4

EFFECTIVE DATE: ..... AUG 5, 1997

SAMPLE TYPE: ..... BEVERAGE ENDS  
FOOD ENDS (CONSULT ALTEK)

NOTE: A SEPARATE SET OF CHANGE PARTS MUST BE MANUFACTURED FOR EACH SPECIFIC DIAMETER OF SAMPLE.

9818: For round, converted/non-converted ends

End Diameter ..... 200 - 307  
( 2" to 3 7/16" )

( For end diameters larger than 307 consult ALTEK )